

PROPOSAL REQUEST

Automated External Defibrillators

For the

**Sumner County
Emergency Medical Service**



**SUMNER COUNTY GOVERNMENT
SUMNER COUNTY, TENNESSEE**

Bid # 20161221-Re-bid

November 2016

Introduction

Sumner County Government is hereby requesting a proposal for Automated External Defibrillators for Sumner County Emergency Medical Service 255 Airport Rd Gallatin, Tennessee 37066.

Sumner County EMS is looking to Purchase 10 Automated External Defibrillators (AED)

Sumner County EMS with this Bid Proposal will Trade In 5 (five) Zoll Plus AEDs.

Total Bid Amount:_____ (Amount MUST include trade-in value of the Zoll AEDs)

Price per AED:_____

Included in the price: AED Adult Pads: yes_____ no_____

AED Pediatric Pads: yes_____ no_____

Child/infant key: yes_____ no_____

Protective Case: yes_____ no_____

Total trade-in value for 5 (five) Zoll AED units:_____

Addition information: *Do not include these into the price of the bid.*

Price of replacement batteries:_____

Price of AED pads:_____

Price of Protective cover/case:_____

Warranty Information:

Are software updates based on AHA Guidelines included free of charge: Yes_____ No:_____

If no what is the cost of the software upgrades:_____

Recommended time frame and cost:

Time Frame:_____ Cost:_____

What is the Manufacture Warranty for Bench Repair? _____

Is a loaner unit available if a unit is sent in for repair? Yes ____ No ____

Is there a cost for the use of a loaner: Yes ____ No ____ Cost: _____

Once Warranty has expired, what is the hourly Factory Service Charge? _____

AED Adult/Pediatric Pads should be compatible with the Philips MRX Heartstart ALS Cardiac Monitor.

Bid must be good for 1 (one) year with the option for a 2 (two) year extension for the purchase of a single unit without trade in:

It is the intent of this bid document to make available to other government entities, by mutual agreement with the successful Bidder, the right to purchase the same or like equipment and or products, at the prices quoted for the period of the contract. Each Bidder shall indicate in the bid proposal, if the Bidder will honor other political subdivision's orders, in accordance with contract terms and conditions.

General Information

I. Proposal Package

All sealed proposal packages must include all of the following, when applicable. Any sealed proposals shall be rejected as a non-conforming bid if any applicable item is missing.

One (1) Original and Two (2) copies.

Evidence of a valid State of Tennessee Business License and/or Sumner County Business License

Evidence of compliance with the Sumner County Government's Insurance Requirements, if work is performed on Sumner County Property

Signed and completed ~~State Bid Form~~ Non-Collusion (Attachment 1)

Properly completed Internal Revenue Service Form W-9

Evidence of a company's safety program and, if supported, a drug testing program (Attachment 2) Drug-Free Workplace

Discussions may be conducted with the vendors which have submitted proposals determined to be reasonably likely of being considered for selection to assure a full understanding of and responsiveness to the RFP requirements. Every effort shall be afforded to assure fair

XX. Applicable Law

Sumner County, Tennessee is an equal opportunity employer. Sumner County does not discriminate towards any individual or business on the basis of race, sex, color, age, religion, national origin, disability or veteran status.

The successful contractor(s) agrees that they shall comply with all local, state, and federal law statutes, rules, and regulations including, but not limited to, the Rehabilitation Act of 1973 and the Americans with Disabilities Act.

In the event that any claims should arise with regards to this contract for a violation of any such local, state, or federal law, statutes, rules, or regulations, the provider will indemnify and hold Sumner County harmless for any damages, including court costs or attorney fees, which might be

5. The County reserves the option to extend the contract term for two (2) additional one (1) year periods. **Yes**_____ **No** _____
6. The winning bidder may submit to the County, for its consideration, a contract price adjustment, not less than sixty (60) calendar days in writing prior to the beginning of each successive term.
Yes_____ **No** _____

All items listed in the bid section will be awarded to only one vendor. The company must have the ability to provide all listed supplies. In the event that the bidder does not have the ability to provide a given item, or a comparable substitute, the bidder is to notate NB in the _____ per uni and the _____ per estimated quantity to be _____ sections next to the applicable item on the Bid Section. Any listed supplies showing (NB-no bid) will be non- conforming.

Bidder understands and accepts the non-appropriation of funds provision of the Sumner

Specifications:

Waveform	Truncated Exponential Biphasic. Waveform parameters adjusted as a function of each impedance.
Therapy	Adult defibrillation peak current: 32A (150J nominal) into a 50 ohm load. Pediatric defibrillation (with optional Infant/Child Key installed): 19A (50J nominal) into a 50 ohm load.
Battery	Standby life (after insertion): Typically 4 years when stored and maintained according to directions provided in Manual.
Protocol	Device Protocol follows preconfigured settings. Defibrillation and CPR protocol can be customized using Configure software.
Quick Shock	Able to deliver a shock after the end of a CPR interval, typically in eight seconds.